

Louisiana Office of Public Health Laboratories	
Test Name	Beta Lactam Inhibitor Tests
PHL Location	Amite Regional Laboratory
CPT Code	N/A
Synonyms	Antibiotic tests: Delvotest 5P and Charm 3 SL3
Brief Description of Test	Delvotest 5P test used for routine screening and confirmation of Beta lactam in raw bovine and goat milks and in Finished Fluid milk products. Charm 3 SL3 is used in testing raw milk samples found positive by industry labs
Possible Results	Positive for the presence of inhibitor or Not Found
Reference Range	Not Found
Specimen Type	Raw dairy milk and fluid finished milk products (cultured products are not tested)
Specimen Container(s):	Raw milk samples are collected in vials and submitted in sterile Whirl-Pak bags or sterile plastic vials normally used by industry. Finished fluid milk products are collected in the container in which they are sold at the retail outlet.
Minimum volume accepted:	Raw milk: at least 40 ml Finished fluid milk products: at least a half pint (when collecting at this volume two samples are required to insure enough volume for all testing)
Collection Instructions	Collect a representative sample according to established protocol. In using the sterile Whirl-Pak or sterile vial there must be at least 0.5 to 1.0 inch air space. If using the Whirl-Pak bag secure closure so that the tabs are not twisted together.
Storage and Transport Instructions	Samples must be stored on wet ice, the ice cannot cover the sample top and there must not be any water over the top of the sample at anytime. The sample must be transported in a timely manner so that testing can be performed prior to 60 hours of collection. Samples must be accompanied with a pilot sample of the same type and size for raw samples and at least ½ the size of the largest container for finished fluid products. The temperature pilot must be from 0.0 - 4.5C upon receipt.
Causes for Rejection	<ol style="list-style-type: none"> <li>1. Temperature of pilot samples out of acceptable range.</li> <li>2. Time of receipt &gt;60 hours from time of collection</li> <li>3. Sample container unprotected or under ice or water when received.</li> <li>4. Sample frozen</li> <li>5. Temperature control missing</li> <li>6. Sample leakage</li> </ol>

	<ul style="list-style-type: none"> <li>7. Unsatisfactory sample container</li> <li>8. Insufficient sample quantity</li> </ul>
Limitations of the Procedure	Delvotest 5P does not specify which Beta lactam or other inhibitor is present and the Charm SL-BL does not specify which Beta lactam drug is present
Interfering Substances	A sample of abnormally high butterfat content.
References	<ul style="list-style-type: none"> <li>Standard Methods for the Examination of Dairy Products</li> <li>FDA 2400 Series for the Examination of Dairy Products</li> <li>US FDA/CFR Grade A Pasteurized Milk Ordinance</li> <li>Laboratory Milk and Dairy Products SOP Manual</li> </ul>
Additional Information	A STARLIMS MDP Collection Manifest or a LAB50 must accompany the submission of samples. All samples and sample information should be pre-logged into the STARLIMS system, including Client Name, product type, container size, name of collector, collection date/time, and temperature of product at collection. A pilot temperature sample must be submitted for each ice chest if multiple ice chests are used.
Release Date	06/12/2019
<p>Warning: If you have printed a copy of this information please be advised that the Louisiana Office of Public Health Laboratories website and methods are updated on a regular basis. Please check the on-line version of this document to ensure you are relying on the most recent release.</p>	