

Name of Agency	
Date	
Surveyor Names:	

Equipment and Supplies	EMT/AEMT/PARAMEDIC & Unit Number								
UNIT #s →									
*Paramedic Level Providers: <i>If the vehicle that is being surveyed makes service runs at either the ALS or BLS level depending on assignment, inspect the vehicle at the highest level used (ALS).</i>									
§6057B.1. TWO-WAY RADIO/DAY TO DAY COMMUNICATION									
Must have a national standard public safety two-way radio communication (day-to-day communications). The ambulance dispatch center(s) and/or point(s) of dispatch must be capable of interactive two-way communications within all of the service's defined area.									
§6057 B.2. TWO-WAY RADIO/DISASTER COMMUNICATION CABILITY:									
Two-way radio with disaster communications capability with the Statewide LWIN system, with the following zones required: a. LERN; b. Bureau of EMS; and c. Statewide Interoperability Channels									
§6057 B.3. DIRECT COMMUNICATION with a PHYSICIAN and HOSPITAL									
Must be conducted through an appropriate system sufficient to ensure adequate communication such as a radio compatible with the Statewide LWIN system; or wireless telephone; or RTSS; or Med. 10 System, etc.									
Act 557. MEDICAL SUPPLIES and EQUIPMENT									
Develop and maintain a written policy identifying the equipment required to conduct, maintain, or operate an ambulance, in fully operational condition, consistent with the scope of practice for emergency medical technicians established in R.S. 40:1133.14.									
The policy identifies the trauma equipment, medications, suction and oxygen equipment, cardiopulmonary resuscitation equipment, and any other equipment required by law that shall be maintained on the ambulance.									

Equipment and Supplies	EMT/AEMT/PARAMEDIC & Unit Number								
UNIT #2 →									
The EMT Level emergency response vehicle equipped in accordance with the written policy.									
The AEMT Level emergency response vehicle equipped in accordance with the written policy.									
The PARAMEDIC Level emergency response vehicle equipped in accordance with the written policy.									
§6057 B.6i. AUTOMATED EXTERNAL DEFIBRILLATION									
All Emergency response vehicles that are not staffed & equipped to the Paramedic level must carry an AED (automatic or semi-automatic) with the appropriate lead cables and at least two sets of the appropriate disposable electrodes. If AED is capable of manual defibrillation, then an appropriate lock out mechanism (such as an access code, computer chip, or lock and key) to prevent unauthorized use of the device by those persons not authorized to manually defibrillate must be an integral part of the device									
INFECTION CONTROL EQUIPMENT									
Infection Control equipment and supplies are readily available for each service run that includes but is not limited to: gloves, face and eye protection/shields, disinfectants and waterless hand cleaners, and sharps containers and biohazard waste trash bags.									
SURVEYOR NOTES:									

Surveyor Signature: _____

Date: _____

(BEMS-AESN revised 07/01/2019)